



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

| APPLICATION NO. | FILING DATE | FIRST NAMED INVENTOR | ATTORNEY DOCKET NO. | CONFIRMATION NO. |
|-----------------|-------------|----------------------|---------------------|------------------|
| 10/714,529      | 11/14/2003  | Alvin Mark Terry     | 41942-05553         | 2211             |

25231 7590 08/23/2005

MARSH, FISCHMANN & BREYFOGLE LLP  
3151 SOUTH VAUGHN WAY  
SUITE 411  
AURORA, CO 80014

EXAMINER

WINAKUR, ERIC FRANK

| ART UNIT | PAPER NUMBER |
|----------|--------------|
|----------|--------------|

3736

DATE MAILED: 08/23/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

|                              |                        |                     |  |
|------------------------------|------------------------|---------------------|--|
| <b>Office Action Summary</b> | <b>Application No.</b> | <b>Applicant(s)</b> |  |
|                              | 10/714,529             | TERRY, ALVIN MARK   |  |
|                              | <b>Examiner</b>        | <b>Art Unit</b>     |  |
|                              | Eric F. Winakur        | 3736                |  |

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☐ Responsive to communication(s) filed on \_\_\_\_.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 53-76 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 53-76 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 14 November 2003 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)                        | 4) <input type="checkbox"/> Interview Summary (PTO-413)                     |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)               | Paper No(s)/Mail Date. ____   |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| Paper No(s)/Mail Date <u>4/22/04</u> .   | 6) <input type="checkbox"/> Other: ____                                     |

## DETAILED ACTION

### ***Claim Rejections - 35 USC § 102***

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 53 - 60, 63, 68 - 73 are rejected under 35 U.S.C. 102(b) as being anticipated by Diab et al. Diab et al. teach an oximeter arrangement (Figure 12; column 15, line 37 - column 16, line 34) for processing red and infrared light samples including performing Fourier transforms on the measured data. The frequency domain information can be further processed to extract the pulse rate information (see Figure 12 elements 1260, 1252 and Figure 17 and the description of "Transform Based Pulse Rate Detection" which begins in column 20). The Fourier transforms can be implemented with fast Fourier transforms (column 22, lines 54 - 59).

### ***Double Patenting***

3. A rejection based on double patenting of the "same invention" type finds its support in the language of 35 U.S.C. 101 which states that "whoever invents or discovers any new and useful process ... may obtain a patent therefor ..." (Emphasis added). Thus, the term "same invention," in this context, means an invention drawn to identical subject matter. See *Miller v. Eagle Mfg. Co.*, 151 U.S. 186 (1894); *In re Ockert*, 245 F.2d 467, 114 USPQ 330 (CCPA 1957); and *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970).

A statutory type (35 U.S.C. 101) double patenting rejection can be overcome by canceling or amending the conflicting claims so they are no longer coextensive in scope. The filing of a terminal disclaimer cannot overcome a double patenting rejection based upon 35 U.S.C. 101.

Art Unit: 3736

4. Claims 65 and 69 are rejected under 35 U.S.C. 101 as claiming the same invention as that of claims 21 and 31 of prior U.S. Patent No. 6,650,918. This is a double patenting rejection.

5. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

6. Claims 53 - 64, 66 - 68, and 70 - 76 are rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1, 2, 4, 6, 13 - 15, 21 - 23, 31, 32, and 35 - 40 of U.S. Patent No. 6,650,918. Although the conflicting claims are not identical, they are not patentably distinct from each other because the claims of the instant application are broader than those of the patent. Thus, any method or apparatus meeting the limitations of the patent would necessarily meet those of the instant application.

### **Conclusion**

7. Applicant cites several references related to optical measurement and determination of pulse rate and constituent levels. However, none of the prior art teaches or suggests a method including computing a value correlated with a blood

Art Unit: 3736


analyte level from DC levels of a first and second plethysmographic signal from a second domain and AC levels from the signals in a first domain, in combination with the other claimed steps. Further, the prior art does not teach or suggest a method of determining a pulse rate that includes obtaining first, second, and third domain estimates of the pulse rate and determining a best estimate of the pulse rate based on at least that first, second, and third domain based estimates, in combination with the other claimed steps. The prior art does not teach or suggest a pulse oximeter that includes a digital processor enabled to construct a second domain filter based on at least the pulse rate of the patient, filter plethysmographic signals in the second domain using the filter, and transform the filtered signals to a first domain, in combination with the other claimed elements.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Eric F. Winakur whose telephone number is 571/272-4736. The examiner can normally be reached on M-Th, 7:30-5; alternate Fri.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eleni Mantis-Mercader can be reached on 571/272-4740. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3736

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Eric F Winakur  
Primary Examiner  
Art Unit 3736

19 August 2005